

# G15 Datasheet



AQL 1.5

EN455 Parts 1, 2, 3 & 4

Class 1 Medical Device

MDD 93/42/EEC



## General Details

Material	PVC plastisol
Surface Finish	Powder Free
Colour	Clear
Geometry	Ambidextrous
Texture	Smooth
Cuff and Finishing	Beaded

## Glove Specifications

Size	Weight g/pc	Length mm	Palm Width mm	Thickness		
				Cuff	Palm	Finger
XS	3.5	230	75 ± 5	0.06 ± 0.02	0.09 ± 0.02	0.11 ± 0.02
S	4	230	85 ± 5	0.06 ± 0.02	0.09 ± 0.02	0.11 ± 0.02
M	4.5	230	95 ± 5	0.06 ± 0.02	0.09 ± 0.02	0.11 ± 0.02
L	5	230	105 ± 5	0.06 ± 0.02	0.09 ± 0.02	0.11 ± 0.02
XL	5.5	230	115 ± 5	0.06 ± 0.02	0.09 ± 0.02	0.11 ± 0.02
Tensile strength before ageing (all gloves)				Load @ peak MPA		Min.13
Tensile strength after ageing (all gloves)				Load @ peak MPA		Min.13

## Performance Requirements

Test	Related Defects	Inspection Level	AQL
Water Leak at Final Release	Holes	G-1 Single Sampling Plan	1.5

## Manufacturing Standards

Maunufacturing Standards	Incompliance with ISO 9001:2000
Presentation	Standard: 100 gloves per dispenser box; 10 dispenser boxes per Carton
Glove Care	Should not store under direct sunlight or at high heat and humid conditions. Single use only. Avoid high polar solvents.



**G·S·L**

WWW.GLOVEMAN.CO.UK

**Gloveman Supplies Ltd**

Vision House, Treleigh Ind Est,

Redruth TR16 4AX

t. 01209 314759

f. 01209 314760

Powder Free Vinyl Exam Gloves